

**Certificate of Compliance/Analysis****Catalog Number:** P40103 Non Sterile.**Product Description:** 100 ml Sample container**Mfg. Lot Number:**1412039**Material:** Vial – PP (Polypropylene),**Cap- PE (Polyethylene)****Material:**

Complies with the Systemic & intracutaneous tests with four different Extractants as per USP Vi. The grade & additives incorporated in it also comply with the FDA: CFR Title 21,177.1520 Olefin polymers. Free from heavy metals according to ELV directive 2000/53/EC and its following amendments.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been found OK for the following characteristics:

**Test/Procedure Result**

Visual attributes	Pass
Dimensional	Pass
Leak test@ 350mm/Hg	Pass
Autoclave test @121°C at 15 psi for 20 minutes.	Pass
Boiling test @ 100°C for 2 Hrs.	Pass
Graduation ( ± 2 ml )	Pass

**Valid****Product:** Best before 5 Yrs from Mfg.**Quality Assurance Incharge****Abdos Labtech Pvt. Ltd.**